

ATP, AMP Analysis

1. Analytical Condition

YL9100 HPLC

Mobile Phase : 0.1M KH₂PO₄(pH=6) aqueous solution : MeOH = 90:10(in one bottle)

Flow Rate : 1ml/min

Column Oven : 30°C

Detector : UV/Vis 254nm

Column : C18 (4.6*250mm, 5µm)

Injection Volume : 20µl

2. Chromatogram

