

Residual Solvents Analysis(Manual Inejction)

1. Analytical Condition

YL 6500 Series

Oven : 35°C(8min)->5°C/min->150°C(10min)

Column : HP-5(30m*0.32mm*1.0um)

Carrier gas : He, 3ml/min (Split ratio 10:1)

Injector : Capillary 230°C

Detector : FID 250°C

Injection Volume : 1ul

2. Chromatogram

