

1,4-Dioxane Analysis

1. Analytical Condition

YL 6500 Series

Oven : 50°C(1min)->10°C/min->200°C(10min)

Column : HP-VOC(60m*0.32mm*1.8um)

Carrier gas : He, 2ml/min (Split ratio 10:1)

Injector : Capillary 230°C

Detector : FID 250°C

Headspace : 100°C, 30min heating, Gas 1ml injection

2. Chromatogram

