

## General Solvents Analysis - THF, MEK, DMAc

### 1. Analytical Condition

#### YL 6500 Series

Oven : 45°C(1min)->10°C/min->200°C(1min)

Column : HP-INNOWAX(30m\*0.53mm\*1.0um)

Carrier gas : He, 8ml/min (Split ratio 5:1)

Injector : Capillary 250°C

Detector : FID 250°C

Sample preconditioning : 80 °C, 20min heating, Gas 1ml injection

### 2. Chromatogram

