

DHA, EPA Analysis In Refined Fish Oil

1. Analytical Condition

YL 6500 GC

Oven : 200°C

Column : HP-INNOWAX(30m*0.53mm*1.0um)

Carrier gas : He, 7.0 ml/min (Split ratio 10:1)

Injector : Capillary 250°C

Detector : FID 250°C

Sample : Refined Fish oil (Preconditioning : FAME)

Injection Volume : 1.0ul (Liquid)

2. Chromatogram

