

Solvent Analysis-EA, Toluene, Xylene

1. Analytical Condition

YL 6500 Series

Oven : 35°C(5min)->5°C/min->200°C(1min)

Column : HP-INNOWAX(30m*0.53mm*1.0um)

Carrier gas : He, 5ml/min (Split ratio 50:1)

Injector : Capillary 250°C

Detector : FID 250°C

Sample preconditioning : 80 °C, 15min heating, Gas 0.3ml injection

2. Chromatogram

