

Glycerol Analysis(On Column)

1. Analytical Condition

YL 6500 Series

Oven : 55°C(1min)->15°C/min->180°C(0.1min)

->7°C/min->230°C(0.1min)-> 30°C/min->370°C(5min)

Column : HT5 (12m*0.53mm*0.15um)

Carrier gas : He, 6ml/min

Injector : On-Column Temp mode : Track Oven

Detector : FID 370°C

Injection Volume : 1ul

2. Chromatogram

